
# Photocurable Nanocomposites for Cartilage Replacements

## **Project Goals**

The properties and characteristics of photocurable nanocomposites can be tailored to mimic those of tissues and/or cartilage, allowing the bioinspired synthetic material to replace them. The goal of this project is to characterize the mechanical properties of this material to better predict its behavior once implemented as a customizable, 3D-printed cartilage replacement.

### **Background and Methods:**


- Previously, nanocomposites with non-functionalized silica (SiO<sub>2</sub>) nanoparticles were characterized.
- In this project, a new subset of functionalized nanoparticles (SiO<sub>2</sub>-Ma) were characterized 0, 3.8, 7.4, 10.7 and 13.8 wt% NPs.



The findings in this project show that the composite generally incurred more integrity allowing it to withstand greater stress and load, thereby demonstrating the tunability of the composites. This project increased the range and resolution to which properties can be selected for 3D-printed cartilage replacements.

I would like to thank Professor Matthew Green for furthering my research opportunities as well as the members of Green Research Group for their time and advice. I would also like to thank the FURI grant for their generous support. This work was partially supported by a NASA ECF Award (80NSSC18K1508) and the Army Research Office (W911NF-18-1-0412).

Alexis Hocken, Chemical Engineering Mentor: Professor Matthew D. Green School of Engineering for Mass, Transport, and Energy



#### Conclusion

#### Acknowledgments

